

Operational Guide for

HBM-1 / HTG-1



May 20th, 2024

Contents




contents


- **Operation**
- **IOL Calculation**
- **Field test results**
- **Myopia management**
- **User Setup**
- **Q&A**

How to activate HBM-1


- ↓ Axial Length(AL) x6
- ↓ Anterior Chamber Depth(ACD)
- ↓ Central Corneal Thickness(CCT)
- ↓ Lens Thickness(LT) x8
- ↓ Zernike Ocular aberrations
- ↓ Pupilometry
- ↓ White to white x1




Unlock packing lock



Turn on Power




Insert ID/PW: huvitz / huvitz




Run the Software


How to activate HBM-1




Check calibration status **regularly**



Place a model eye on the chinrest




Check the message "Data validated"



Set up the focus like the picture

Workflow



Operation

Patient Info

New Patient

Patient ID: 00032 | First Name: Sook | Last Name: Choi

Birth: 1984-09-23 | Gender: Female | Physician: 1

Cancel OK

00005	Patient_Keratoconus2	Male	1983-08-17	2023-08-17
00004	Patient_Keratoconus1	Male	1983-08-17	2023-08-17

Measurement

MEASURE

OD Phakic | MEASURE | OS Phakic

AL	24.56 mm	24.31 mm
CCT	0.487 mm	0.490 mm
ACD	3.11 mm	3.12 mm
LT	3.89 mm	3.92 mm
K1	8.76 mm	8.73 mm
K2	8.66 mm	8.63 mm
AXIS	24°	18°

BIOMETER | PUPIL

✓ KER | ✓ AL | ✓ ANT

Result

RESULT

OD Phakic | OS Phakic

AL (mm)	24.26	24.37
CCT (mm)	0.547	0.538
ACD (mm)	3.11	3.13
LT (mm)	3.61	3.64
K1 (mm)	8.15	8.18
K2 (mm)	7.94	7.99
AXIS (°)	177	6

Report Preview

Report Preview showing data for Patient 2, including OD and OS measurements, keratometry, and topography. The report includes a table of data measurements and topography images.

Workflow

1. Easy & fast workflow
2. Auto tracking & shoot
3. Separate measurement

Patient Page

Operation



Axial Length(AL) x6
 Anterior Chamber Depth(ACD)
 Central Corneal Thickness(CCT)
 Lens Thickness(LT)
 X2
 K1
 Te
 K2 at C
 Zernike Co
 Pupilometry
 White to white
 x1

HBM-1 PATIENT LIST

Searching Today List New Delete Transfer (HIIS-1)

Search [Q] [Calendar] [Person+] [Trash] [Filter] [Menu]

ID	Name	Gender	Age	Update
00014	김재민	Male	1983-10-13	2023-10-13
00013	임인호	Male	1983-10-13	2023-10-13
00012	김정환	Male	1983-10-13	2023-10-13
00011	이화준	Male	1983-10-13	2023-10-13
00010	송인석	Male	1983-10-13	2023-10-13
00032	Sook Choi	Female	1984-09-23	2023-10-13
00002	Patient_2	Male	1983-08-17	2023-10-12
00001	Patient_1_myopia	Male	2006-08-17	2023-10-05
00007	Patient DCM_IOL	Male	1983-08-17	2023-09-26

[Person] Current Selection
 [Camera] Disabled
 [Eye] User Option
 [Bar Chart] Power Off

New Patient

Operation



New Patient

Patient ID

00032

First Name

Sook

Last Name

Choi

Birth YYYY-MM-DD

1984-09-23

Gender

Female

Physician

1

Each Physician can have its own lens & formulas

Cancel

OK

00005	Patient_Keratoconus2	Male	1983-08-17	2023-08-17
00004	Patient_Keratoconus1	Male	1983-08-17	2023-08-17

Patient Detail Page

Operation



HBM-1 **PATIENT LIST**

← Patient ID 00011 Gender Male Physician
Name 이화준 Age 1983-10-13

MYOPIA

Start Myopia management

Delete Patient

DATE / TIME	AL	CCT	ACD	LT	K1	K2	Axis	Lens
OD 2023-10-13	24.82	0.594	3.10	4.76	7.99	7.64	179	Phakic
OS 14:48:05	24.18	0.594	3.06	4.75	7.88	7.60	0	Phakic

Press this button to start a measurement

Measurement Page

Operation



MEASURE

OD Phakic

MEASURE

OS Phakic

LENS

- Phakic
- Aphakic
- Pseudophakic (Unknown)
- Pseudophakic (Silicone)
- Pseudophakic (PMMA)
- Pseudophakic (Acrylate)
- Pseudophakic (Memory)
- Myopia

Spherical

START

Keratometry Focusing
Open your eyes comfortably

Select a crystal lens type

Phakic - Include crystalline lens
Aphakic - Exclude crystalline lens
Pseudophakic - IOL surgical eye

Axial Length
(The method of calculating Axial Length depends on the type of lens)
Phakic : + 0.0mm
Aphakic : + 0.20mm
Pseudophakic : + 0.11 ~ 0.12mm

Input the Diopter(Optional)

START

KER AL ANT

Measurement

Operation



Right Auto Shoot Left

MEASURE

Auto Tracking Keratometry Focusing Open your eyes comfortably Live Signal Clear Retry

Message window

Move HBM body to the center roughly

Selectable

BIOMETER PUPIL

KER AL ANT

Each mode can be on or off

Measurement Page

OD Phakic AL mm STD CCT mm STD ACD mm STD LT mm STD K1 mm K2 mm AXIS °

OS Phakic CCT mm STD ACD mm STD LT mm STD K1 mm K2 mm AXIS °

Sook Choi

Measurement

Operation



HBM-1 MEASURE kevin 00008

OD Phakic OS Phakic

MEASURE

AL 24.96 mm 0.039 STD
24.96

CCT 0.495 mm 0.020 STD
0.487

ACD 3.58 mm 0.020 STD
3.58

LT 3.73 mm 0.010 STD
3.74

K2 44.71 D

AXIS 126 °
Cyl = -0.51D

AL 24.83 mm 0.014 STD
24.83

CCT 0.482 mm 0.020 STD
0.482

ACD 3.60 mm 0.020 STD
3.60

LT 3.73 mm 0.020 STD
3.73

K1 44.14 D

K2 45.65 D

AXIS 173 °
Cyl = -1.51D

Keratomy Focusing
Open your eyes comfortably

Green : Data is passed

Orange : Data is suspicious

Red : Data is failed

BIOMETER PUPIL

If 'Auto Try' is on, process 3 times until data is satisfied

✓ KER ✓ AL ✓ ANT

Go to result page

Measurement(Bad Case Ker Tear 1)

Operation



HBM-1

RESULT



Cristea Elena
F2-00002

2023-11-27
18:27:01

OD

OS

MAP

PROFILE

SimK

K1(Rf)

44.34D(7.61mm) @81°

K2(Rs)

45.75D(7.38mm) @171°

CYL

-1.41D

K Max(R Min)

45.87D(7.36mm) @3°

K Avg

45.04D(7.49mm)

Ecc(8mm)

e

0.61

SUM

KER

ZER

AL

ANT

PUPIL

WTW

CONT.FIT



EYE

MAP

RING

Axial

SCALE

Measurement(Bad Case Ker Tear 2)

Operation



HBM-1

RESULT



Muresan Ana
F2-00007

2023-12-04
17:30:49

OD

OS

MAP

PROFILE

SUM

KER

ZER



AL

ANT

PUPIL

WTW

CONT.FIT



SimK

K1(Rf)

45.52D(7.41mm) @23°

K2(Rs)

46.81D(7.21mm) @113°

CYL

-1.29D

K Max(R Min)

48.65D(6.94mm) @84°

K Avg

46.16D(7.31mm)

Ecc(8mm)

e

0.30

EYE

MAP

RING

Axial

SCALE

Measurement(Bad Case Ker Tear 3)

Operation



HBM-1

RESULT



Petrican Aurica
F2-00012

2023-12-06
17:31:28

OD

OS

MAP

PROFILE

SimK

SUM

KER

ZER


AL

ANT

PUPIL

WTW

CONT.FIT



K1(Rf)

43.82D(7.70mm) @91°

K2(Rs)

44.98D(7.50mm) @1°

CYL

-1.16D

K Max(R Min)

46.27D(7.29mm) @176°

K Avg

44.39D(7.60mm)

Ecc(8mm)

e

-0.58

EYE

MAP

RING

Axial

SCALE

Measurement(Bad Case Ker Eyebrow)

Operation



HBM-1 **RESULT** Ye J 40178 2023-02-28 15:49:24

OD **OS** MAP PROFILE

SUM **KER** ZER AL ANT PUPIL

54.00
53.00
52.00
51.00
50.00
49.00
48.00
47.00
46.00
45.00
44.00
43.00
42.00
41.00
40.00
39.00
38.00
37.00
36.00
35.00
34.00
33.00
32.00
31.00

SimK

K1(Rf) 41.95D(8.04mm) @2°

K2(Rs) 43.44D(7.77mm) @92°

CYL -1.49D

K Max(R Min) 44.66D(7.56mm) @93°

K Avg 42.68D(7.91mm)

Ecc(8mm) -0.90

SETUP

System Auto Tracking **On** Off Auto Retry

Patient Auto Shoot **On** Off

Measure **KER** **AL** **ANT**

! We recommend manual Ker measurement by operator !

Result : Summary

Operation



HBM-1
RESULT

kevin
00008

2024-04-12
13:59:18

	OD	OS
AL (mm)	24.97 <small>0.036 STD</small>	24.82 <small>0.032 STD</small>
CCT (mm)	0.492	0.455 <small>0.010 STD</small>
ACD (mm)	3.59	3.63 <small>0.020 STD</small>
LT (mm)	3.70 <small>0.050 STD</small>	3.75 <small>0.020 STD</small>
K1	44.27	
WTW		
AXIS	142	

Warning Mark (!)

AL Warning - Data was obtained by DCM Mode

ACD Warning - Patient Lens is Aphakic / PseudoPhakic

LT Warning - Patient Lens is Aphakic / PseudoPhakic

Result : Keratometry

Operation



HBM-1
RESULT
00032
2024-02-14 15:45:21

OD

OS

SUM

KER

ZER

AL

ANT

PUPIL

WTW

CONT.FIT

MAP

PROFILE

EYE

MAP

RING

Axial

SCALE

SimK

K1(Rf) 38.82D(8.69mm) @6°

K2(Rs) 39.45D(8.55mm) @96°

CYL -0.63D

K Max(R Min) 39.28D(8.59mm) @127°

K Avg 39.13D(8.62mm)

-0.94

Zoom Image

Eye Image ON/OFF

Map Image ON/OFF

Placido Ring Segmentation ON/OFF


Axial, Tangential, Elevation, Refractive

Select Color palette

Select Overlay

Result : Topography Zoom Image

appendix



Result : Topography 4Maps

appendix







Display Type
4 Images

Scale


Ker Data

Rf	8.69mm
K1	38.82D
Axis(K1)	63°
Rs	8.64mm
K2	39.07D
Axis(K2)	153°
Cyl	-0.25D
R Min	8.62mm
K Max	39.17D
R Avg	8.67mm
K Avg	38.94D
Ecc(8mm)	-0.90(e)




Result : Topography Comparison

appendix



Result



Operation

HBM-1 **RESULT** 00032 2024-02-14 15:45:21

SimK

K1(Rf) 38.82D(8.69mm) @6°

K2(Rs) 39.45D(8.55mm) @96°

CYL -0.63D

K Max(R Min) 39.28D(8.59mm) @127°

K Avg 39.13D(8.62mm)

Ecc(8mm) e -0.94

Meridian

2mm Zone

K1 8.78mm @ 12°
K2 8.60mm @ 102°
AVG 8.69mm
Cyl -0.82D ax 12°

4mm Zone

K1 8.63mm @ 4°
K2 8.50 @ 94°
AVG 8.57mm
Cyl -0.58D ax 4°

6mm Zone

K1 8.38mm @ 178°
K2 8.13mm @ 88°
AVG 8.26mm
Cyl -1.25D ax 178°

2mm / 4mm / 6mm
3mm / 5mm / 7mm
Diopter / Millimeter
Changeable

Keratoconus

KPI (Keratoconus Prediction Index) 0.30

SAI (Surface Asymmetry Index) 1.06

DSI (Differential Sector Index) 3.47

CSI (Central/Surrounding Index) 3.47


IAI (Irregular Astigmatism Index) 1.14

0.23

If the KPI is over 0.23,
It suspicious Keratoconus

Central steepening Keratoconus
Keratoconus Suspicious
Non-Keratoconus

Result



Operation

HBM-1

RESULT



Keratoconus Sample
00004

2023-04-27
12:08:18

OD

OS

SUM

KER

ZER

AL

ANT

PUPIL


WTW

CONT.FIT

MAP

PROFILE

Data Profiler



Steepest Meridian

Change Steep Axis



99



Flattest Meridian

Change Flat Axis



9



K-Value (Diopter)

Value



-6.0 -5.0 -4.0 -3.0 -2.0 -1.0 0.0 1.0 2.0 3.0 4.0 5.0 6.0

279°/189°

99°/9°

Differential



-6.0 -5.0 -4.0 -3.0 -2.0 -1.0 0.0 1.0 2.0 3.0 4.0 5.0 6.0

Result: Zernike



Operation

HBM-1

RESULT



Test3 Test3
F1-00028

2023-12-05
11:13:02

OD

OS

SUM

KER

ZER

AL

ANT

PUPIL

WTW

CONT.FIT

MAP

PSF

Low Order Aberrations

High Order Aberrations

Expansion Coefficients


7.0mm



RMS = 0.37 μm


Related with SPH, CYL

7.0mm



RMS = 0.16 μm

- Irregular astigmatism
- Corneal imbalance
- Size of pupil



Result: Zernike



Operation

HBM-1 **RESULT** **Keratoconus Sample 00004** 2023-04-27 12:08:18

OD **OS**

SUM

KER

ZER

AL

ANT

PUPIL

WTW

CONT.FIT

MAP **PSF**


Point Spread Function

Object

Retina

E # * [Small E grid]

Result



Operation

HBM-1
RESULT

송인석
00010
2023-10-13
14:46:27

OD OS

SUM

KER

ZER

AL

ANT

PUP

WTW

CONT.FIT

DCM

Axial Length 26.87 mm
0.022 STD

No.	AL	Status
1	26.89	included
2	26.89	
3	26.83	
4	26.93	excluded
5	26.80	
6	26.89	
AVG		26.90

#1 #2 #3

#4 #5 #6


Axial Length 26.87 mm
0.022 STD

#1 #2 #3

#4 #5 #6

No of Exam

Result



Operation

HBM-1
RESULT

송인석
00010
2023-10-13
14:46:27

OD OS

SUM

KER

SETUP

- System
- Patient
- Measure
- Connectivity
- Report
- IOL
- Information

Language: English

Device Name: _____

Sleep Time: Off 5 Min

Touch Keyboard: On Off

Login Page: On Off

Save Mode: Light **Raw**

DCM

No.	AL
1	26.89
2	26.89
3	26.83
4	26.93
5	26.80
6	26.89
7	26.90

26.87 mm
0.022 STD


#1 #2 #3
#4 #5 #6

Dense Cataract Mode
Analyze data with DCM algorithm

Caution !
If Setup - Save Mode : Light Mode
DCM mode not working
(Light Mode : Don't Save Raw data)

Caution !
Severe cataract may not come out on DCM Mode, which requires the use of A-Scan Ultrasonic Probe

Result



Operation

HBM-1

RESULT



00032

2024-02-14
15:45:21

OD

OS

SUM

KER

ZER

AL

ANT

PUPIL


WTW

CONT.FIT

CCT

ACD

LT



Total AVG

CCT 0.450 mm
0.010 STD

ACD 3.48 mm
0.020 STD

LT 3.65 mm
0.030 STD

No.	CCT	ACD	LT
1	0.459	3.47	3.66
2	0.461	3.47	3.65
3	0.440	3.45	3.68
4	0.459	3.45	3.59
5	0.448	3.46	3.64
6	0.454	3.47	3.62
7	0.429	3.44	3.68
8	0.458	3.50	3.68

AVG 0.451 3.47 3.65

#1

#2

#3


#4

#5

#6

Result : Pupilometry

Operation



HBM-1 **RESULT** Sook Choi 00032 2023-10-13 08:28:51

OD **OS** **PUPIL** **GRAPH** **VALUES**

SUM **Dynamic** Average Diameter

KER Dynamic : Photopic + Mesopic Min Values 2.17mm

ZER Photopic : Daylight Max Values 5.13mm

AL Mesopic : Dark night Average 4.08mm

ANT No / Total Frame Std 0.789mm

PUP Diameter

WTW Diameter 4.53mm

CONT.FIT Pupil Center Diff

X -0.01mm

Y -0.01mm

Time

Time 4.90

R.Center **Pupil** **Grid** **Rulers**

Result : Pupilometry

Operation



Result : Pupilometry

Operation

HBM-1

RESULT

Sook Choi
00032

2023-10-13
08:28:51

OD OS

SUM

KER

ZER

AL

ANT

PUP

WTW

CONT.FIT

PUPIL GRAPH

Average Diameter

Min Values	2.17mm
Max Values	5.13mm
Average	4.08mm

Average Pupil Decentration

X	-0.24mm
Y	-0.17mm
Std Dev	0.12mm

Latency Decentration

Result

- Axial Length(AL) x6
- Anterior Chamber Depth(ACD)
- Central Corneal Thickness(CCT)
- Lens Thickness(LT) x8
- Keratometry
- Topography
- Keratoconus
- Zernike Coefficients
- Pupilometry
- White to white x1



Operation

HBM-1

RESULT



Keratoconus Sample
00004

2023-04-27
12:08:18

OD

OS

Editable

EDIT

VALUES

Diameter

10.69mm


Decentration

X

0.61mm

Y


-1.52mm



Result: WhitetoWhite

Operation

- Axial Length(AL)
- x6
- Anterior Chamber Depth(ACD)
- Central Corneal Thickness(CCT)
- Lens Thickness(LT)
- ...
- Zernike Oculaberrations
- Pupillometry
- White to white



Centering for Y position

Adjustable X Start Position

Move button

Adjustable X End Position

Result: Contact Lens Fitting

Operation



HBM-1

RESULT


Myopia Management Sample 00001 2024-02-01 20:47:34

OD OS

Simulation Gallery

- SUM
- KER
- ZER
- AL
- ANT
- PUPIL
- WTW
- CONT.FIT

- 0
- 75
- 125
- 175
- 225
- 275
- 325
- 375
- 425
- 525
- 575
- 625
- 675
- 725
- 775
- 825
- 875
- 925
- 975
- 1000



Patient SimK

Axial / Tangential Map ON/OFF

Lens Fitting Map ON

Axial

Map

Manufacturer Please Select

Lens Please Select




Set Parameter

Keratometry
K1 42.48 @ 5°
K2 43.15 @ 95°
Cyl -0.67 ax 5°

r0 7.85
Diameter 10.00
dr0 0.00
Ecc 0.30
r1 9.35
Width 1.00
dr1 0.00
Ecc 0.00

Lens Info



SAVE

Result: Contact Lens Fitting

Operation



HBM-1

Lens Parameters

Patient 2

2023-07-19

8:12

Base Curve

Curvature Radius(r0)

- 7.99 +

Diameter(mm)

- 10 +

Curvature Diff(dr0)

- 0 +

Eccentricity

- 0.3 +

Set a base curve

Peripheral Curve 1

Curvature Radius(r1)

- 9.49 +

Width(mm)

- 1 +

Curvature Diff(dr1)

- 0 +

Eccentricity

- 0 +

Set the first peripheral

Peripheral Curve 2

Curvature Radius(r2)

- 0 +

Width(mm)

- 0 +

Curvature Diff(dr2)

- 0 +

Eccentricity

- 0 +

Set the second peripheral

Peripheral Curve 3

Curvature Radius(r3)

- 0 +

Set Parameter

- 0 +

Curvature Diff(dr3)

- 0 +

Eccentricity

- 0 +

Set the third peripheral

Set a proper base and first curve

1

Default

Cancel

2

Apply

Result: Contact Lens Fitting

Operation




RESULT


Patient_2
00002

2023-07-19
15:18:12

OD
OS

SUM

KER

ZER

AL

ANT

PUP

WTW

CONT.FIT

Simulation
Gallery

0

75

125

175

225

275

325

375

425

525

575

625

675

725

775


825

875

925

975


1000



Simulated with a selected lens

Axial

Map



Manufacturer

JenaLens
▼

Lens

Asph
▲

ABOGDT
7°

ABOGVTP
7°

ABOGTpri
7°

Asph
8.05

Asph-Kera
9.80

AsphPri
0.00

Asph VTP
2.00

Jeclips
0.40


Jecl. RT/BT
0.00

Jecl. VT
0.60

< 1/8 >

Result: Print Option

Operation



RESULT

OD OS

SUM

KER

ZER

AL

ANT

PUPIL

WTW

CONT.FIT


Transfer Data

Choose Where To Send Your Data.


	HIIS-1	Dicom	Print
	Send HIIS-1	Send DICOM	Send Printer
Biometry	OFF	OFF	OD/OS/OU/OFF
Topography	OFF	OFF	OD/OS/OU/OFF
Topo Map	Select	Select	Axial / Tangential / Ref Power / Elevation
IOL	OU	OU	OD/OS/OU/OFF
IOL List	IOL C.	Post R	Toric
			OFF
			Barrett Tab

Cancel OK

024-02-01 10:47:34



IOL



IOL Calculation

HBM-1

IOL CALCULATION

Myopia Management Sample
00001
2024-02-13
10:05:57

OD

OS

IOL Calculation

Post Refractive

Toric

Barrett

Reset

Select OD / OS

General IOL

Post Surgery

Toric IOL

Select Surgeon

0

Warning (!)
Measurement - Manual
AL - DCM
ACD,LT - Aphakic /
Pseudophakic

(Optional) Include
Barrett Universal II
Barrett Universal II Toric
Barrett True K
Barrett True K Toric
Barrett Rx
(License Only)

Measurement	Value	Surgeon	Target	AMO	Tecnis 1 ZCB00	SRK2	AMO	Tecnis 1 ZCB00	SRK2	Tecnis 1 ZCB00	SRK2
AL (!)	25.91										
CCT	0.447										
ACD (!)	3.54										
LT (!)	3.70										
K1	38.73										
K2	39.16	A = 119.600	19.28	A = 119.600	19.28	A = 119.600	19.28	A = 119.600	19.28	A = 119.600	19.28
Cyl	-0.44										
Axis	170										
WTW	11.83										
		IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)
		18.50	0.63	18.50	0.63	18.50	0.63	18.50	0.63	18.50	0.63
		19.00	0.23	19.00	0.23	19.00	0.23	19.00	0.23	19.00	0.23
		19.50	-0.17	19.50	-0.17	19.50	-0.17	19.50	-0.17	19.50	-0.17
		20.00	-0.57	20.00	-0.57	20.00	-0.57	20.00	-0.57	20.00	-0.57
		20.50	-0.97	20.50	-0.97	20.50	-0.97	20.50	-0.97	20.50	-0.97

Edit Measurement

Print

IOL : IOL Measurement Data Edit

IOL Calculation

HBM-1 IOL CALCULATION Patient_2 2023-07-19
00002 15:18:12

OD OS IOL Calculation Post Refractive Toric

Measurement Data Edit

Status: Measured

Measured Manual

Biometry			Keratometry							
AL	24.26	mm	CCT	0.547	mm	K1	8.15	mm	Index	1.3375
ACD	3.11	mm	WTW	12.28	mm	K2	7.94	mm	Unit	mm
LT	3.61	mm								

All data is editable in 'Manual' status

Reset Cancel OK

20.50	0.34	21.00	0.33	20.50	0.43	20.00	0.51
21.00	0.00	21.50	-0.02	21.00	0.09	20.50	0.16
21.50	-0.34	22.00	-0.37	21.50	-0.24	21.00	-0.20
22.00	-0.69	22.50	-0.72	22.00	-0.58	21.50	-0.57

IOL : General IOL

IOL Calculation

HBM-1
IOL CALCULATION
Myopia Management Sample 00001
2024-02-13 10:05:57

OD OS
IOL Calculation Post Refractive Toric Barrett
Reset

Surgeon: 2 Target: -1.0 Input target diopter after surgery

Manufacturer: AMO Lens Model: Tecnis 1 ZCB00

Formula: Holladay Barrett Universal II HofferQ SRK2

Selected Lens constants / Ideal IOL Diopter

License Only

The closest IOL diopter to the input target

	IOL(D)	REF(D)	IOL(D)	REF(D)
Barrett Universal II	21.00	-0.96	21.00	-0.97
	21.50	-1.32	21.50	-1.35
	22.00	-1.69	22.00	-1.73

IOL : IOL Post Ref

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample
00001
2024-02-13
10:05:57

OD
OS

IOL Calculation
Post Refractive
Toric
Barrett

Next Page
Reset

Surgeon
LT

Select Surgeon v
3.68

Keratorefractive Surgery Information

Correction Type

Myopic
v

Type

PRK
v

Myopic
Hyperopic

Unknown
RK
PRK
LASIK
LASEK

Camellin Calossi

Shamma's No History

Select formulas to be used

Formulas Available

Camellin Calossi
SIRC

SIRC (Surgically Induced Refractive Change)
"Rx Change"
Example :
 Before refractive surgery : SPH -4D CYL -2 ax 180 -> Preop SEQ = -5D
 After refractive surgery : SPH -0D CYL -1 ax 180 -> Postop SEQ = -0.5D
 SIRC = (-5D) - (-0.5D) = -4.5D

Measurement (!)

AL 25.91

CCT 0.447

ACD 3.54

LT 3.70

K1 38.73

K2 39.16

Cyl -0.44

Axis 170

WTW 11.83

Surgeon
LT

Select Surgeon v
3.68

Keratorefractive Surgery Information

Correction Type

Myopic
v

Type

PRK
v

Myopic
Hyperopic

Unknown
RK
PRK
LASIK
LASEK

Camellin Calossi

Shamma's No History

Select formulas to be used

Formulas Available

Camellin Calossi
SIRC

SIRC (Surgically Induced Refractive Change)
"Rx Change"
Example :
 Before refractive surgery : SPH -4D CYL -2 ax 180 -> Preop SEQ = -5D
 After refractive surgery : SPH -0D CYL -1 ax 180 -> Postop SEQ = -0.5D
 SIRC = (-5D) - (-0.5D) = -4.5D

IOL : IOL Post Ref

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample
00001
2024-02-13
10:05:57

OD
OS

IOL Calculation
Post Refractive
Toric
Barrett

Reset

Measurement (i)

AL 25.91

CCT 0.447

ACD 3.54

LT 3.70

K1 38.73

K2 39.16

Cyl -0.44

Axis 170

WTW 11.83

Surgeon	Target				
2	-1.0	Input target diopter after surgery			
AMO	AMO	AMO	AMO	AMO	AMO
ReZoom NXG1	n 911A	808C	808C	808C	808C
Camellin-Calossi	Camellin-Calossi	Shammas No Hist...	Camellin-Calossi	Camellin-Calossi	Camellin-Calossi
Formula					
A = 118.400	22.51	A = 118.300	22.39	A = 118.000	22.10
A = 118.000	22.10	A = 118.000	22.04	A = 118.000	22.04
IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)
21.50	-0.28	21.50	-0.57	21.00	-0.13
22.00	-0.64	22.00	-0.72	22.00	-0.26
22.50	-0.99	22.50	-1.08	22.00	-0.92
23.00	-1.34	23.00	-1.43	22.50	-1.32
23.50	-1.70	23.50	-1.79	23.00	-1.73

Selected Lens lens constants / Ideal IOL Diopter

The closest IOL diopter to input target

Result: IOL Toric

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample
00001
2024-02-13
10:05:57

OD

OS

IOL Calculation

Post Refractive

Toric

Barrett

>
↶

Surgeon	Target	SIA	IL	
1	-1	1	35	

HumanOptics	Manufacturer	Optics	HumanOptics	HumanOptics
TORICA -aAY Stan...	Lens model	TORICA -aAY Stan...	TORICA -aAY Stan...	TORICA -aAY Stan...
Barrett Univ. II Toric	Formula	SRK/T	SRK/T	SRK/T

SRK/T	A = 118.700	20.04	A = 118.700	20.04
HofferQ				
Haigis				
Holladay				
Barrett Univ. II Toric	Licensed only	IOL(D)	REF(D)	Selected Lens lens constants / Ideal IOL Diopter
		20.50	-0.86	20.00 -0.97

Select a lens

✎

🖨

Result: IOL Toric (Select Target REF)

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample
00001
2024-02-13
10:05:57

OD

OS

IOL Calculation

Post Refractive

Toric

Barrett

>

↶

Surgeon	Target	SIA	IL
1	-1	1	35

HumanOptics	TORICA -aAY Stan...	TORICA -aAY Stan...	TORICA -aAY Stan...	TORICA -aAY Stan...
Barrett Univ. II Toric	SRK/T	SRK/T	SRK/T	SRK/T

A	IOL(D)	REF(D)	A	IOL(D)	REF(D)	A	IOL(D)	REF(D)
118.700	20.04	20.04	118.700	20.04	20.04	118.700	20.04	20.04
	IOL(D)	REF(D)		IOL(D)	REF(D)		IOL(D)	REF(D)
	19.00	-0.22		19.00	-0.22		19.00	-0.22
	20.00	-0.48		20.00	-0.59		20.00	-0.59
	20.50	-0.86		20.00	-0.97		20.00	-0.97
	21.00	-1.24		20.50	-1.35		20.50	-1.35
	21.50	-1.63		21.00	-1.73		21.00	-1.73

Select a lens

When selected IOL REF diopter , next button is available

Input target diopter after surgery

Input SIA (Surgically induced astigmatism)

Input IL (Incision on axis)

A = 118.700 20.04 A = 118.700 20.04 A = 118.700 20.04

- SRK/T
- HofferQ
- Haigis
- Holladay
- Barrett Univ. II Toric

20.50 -0.86 20.00 -0.97 20.00 -0.97 20.00 -0.97

Result: IOL Toric (Select Target CYL)

IOL Calculation



HBM-1 IOL CALCULATION Myopia Management Sample 00001 2024-02-13 10:05:57

OD OS IOL Calculation Post Refractive **Toric** Barrett

Formula: Barrett Univ. II Toric Target: -1 SIA: 1 IL: 35

Previous Page (Set REF Power) Reset

Ideal Cylindrical Power

IOL Ideal Toricity 1.16

Measurement

AL	25.91
CCT	0.447
ACD	3.54
LT	3.70
K1	38.73
K2	39.16
Cyl	-0.44
Axis	170
WTW	11.83

Toric Lense Information

Lens Axis

Available Toric Lenses

Lens	Res Astigm
n.a.	n.a.
Non Toric	The closest IOL toricity to the ideal toricity
TORICA -aA Special	-0.14D @ 33°
TORICA -aA Special	-0.22D @ 123°
TORICA -aA Special	-0.59D @ 123°

Select a lens

Model: TORICA -aA Special

Spherical Equivalent Power (D)	Cylindrical Power (D)
20.50 Selected REF power	1.00 Selected CYL Power

Spherical Power (D)	Axis Of Placement (°)	Expected Refraction
20.00 Selected CYL Power	123 Lens Axis	-0.79D -0.14D @ 33°

Expected result

IOL: IOL (Barrett)



IOL Calculation

HBM-1
IOL CALCULATION
Myopia Management Sample 00001
2024-02-13 10:05:57

OD
OS

Measurement (i)

AL 25.91

CCT 0.447

ACD 3.54

LT 3.70

K1 38.73

K2 39.16

Cyl -0.44

Axis 170

WTW 11.83

Surgeon 2

Target 1

AMO v

Tecnis 1 ZCB00 v

AMO v

Tecnis 1 ZCB00 v

AMO v

Tecnis 1 ZCB00 v

LF = 2.041		18.44		LF = 2.041		18.44		LF = 2.041		18.44	
IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)
17.50	1.64	17.50	1.64	17.50	1.64	17.50	1.64	17.50	1.64	17.50	1.64
18.00	1.30	18.00	1.30	18.00	1.30	18.00	1.30	18.00	1.30	18.00	1.30
18.50	0.95	18.50	0.95	18.50	0.95	18.50	0.95	18.50	0.95	18.50	0.95
19.00	0.60	19.00	0.60	19.00	0.60	19.00	0.60	19.00	0.60	19.00	0.60
19.50	0.25	19.50	0.25	19.50	0.25	19.50	0.25	19.50	0.25	19.50	0.25

Barrett Formulas

Universal II ^

Universal II

Universal II Toric

True K

True K Toric

RX

Formula

Universal II - General IOL

Universal II Toric - Toric IOL

True K - Post Ref IOL

True K - Post Ref Toric IOL

Rx - Exchange IOL / Piggyback IOL

↩

IOL: IOL (Barrett True K / True K Toric)

IOL Calculation

HBM-1 IOL CALCULATION Myopia Management Sample 00001 2024-02-13 10:05:57

OD OS IOL Calculation Post Refractive Toric Barrett

Surgeon 2 Barrett Formulas True K

Target -1 **Input target diopter after surgery** History Myopic Lasik Pre-Lasik Ref -5.0 Post-Lasik Ref -1.0

AL 25.91 CCT 0.447 ACD 3.54 LT 3.70 K1 38.73 K2 39.16 Cyl -0.44 Axis 170 WTW 11.83

AMO AMO CeeOn 911A Tecnis 1 ZCB00

LF = 1.726 TK LF = 1.880 21.60 LF = 2.041 TK LF = 2.200 22.15

TRUE K : 38.5 / Corr. -3.73 D

AMO Tecnis 1 ZCB00

LF = 2.041 TK LF = 2.200 LF = 2.041 TK LF = 2.200

IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)
21.00	-0.55	21.50	-0.52	22.00	-0.89	22.50	-1.26
21.50	-0.93	22.00	-0.89	22.50	-1.26	23.00	-1.64
22.00	-1.31	22.50	-1.26	23.00	-1.64	23.00	-1.64
22.50	-1.69	23.00	-1.64	23.00	-1.64	23.00	-1.64

Post Surgery history

Input Pre Surgery / Post Surgery Ref diopter when nohistory ref diopter
 If input is 'Blank', Correction diopter is calculated by measurement data
 TRUE K - Be used in formula
 It changed depending on the surgery type

IOL: IOL (Barrett True K / True K Toric)

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample
00001
2024-02-13
10:05:57

OD

OS

IOL Calculation

Post Refractive

Toric

Barrett

>

↶

Surgeon 2

Target -1

AMO

CeeOn 911A

AL 25.91

CCT 0.447

ACD 3.54

LT 3.70

K1 38.73

K2 39.16

Cyl -0.44

Axis 170

WTW 11.83

Barrett Formulas True K

History Myopic Lasik

Pre-Lasik Ref -5.0

Post-Lasik Ref -1.0

TRUE K : 38.5 / Corr. -3.73 D

AMO

Tecnis 1 ZCB00

LF = 1.726	TK LF = 1.880	21.60	LF = 2.041	TK LF = 2.200	22.15	LF = 2.041	TK LF = 2.200	22.15	LF = 2.041	TK LF = 2.200	22.15
IOL(D)	REF(D)		IOL(D)	REF(D)		IOL(D)	REF(D)		IOL(D)	REF(D)	
20.50	-0.18		21.00	-0.16		21.00	-0.16		21.00	-0.16	
21.00	-0.55		21.50	-0.52		21.50	-0.52		21.50	-0.52	
21.50	-0.93		22.00	-0.89		22.00	-0.89		22.00	-0.89	
22.00	-1.31		22.50	-1.26		22.50	-1.26		22.50	-1.26	
22.50	-1.69		23.00	-1.64		23.00	-1.64		23.00	-1.64	

Input target diopter after surgery

- Myopic Lasik
- Myopic Lasik
- Hyperopic Lasik
- Radial Keratotomy
- Keratoconus

Selected Lens lens constants / Ideal IOL Diopter

The closest IOL diopter to the input target

IOL: IOL (Barrett True K / True K Toric)

IOL Calculation



HBM-1
IOL CALCULATION
Myopia Management Sample 00001
2024-02-13 10:05:57

OD

OS

IOL Calculation

Post Refractive

Toric

Barrett

>

↶

Surgeon 2 Barrett Formulas True K Toric

Target -1 SIA 1 IL 45 History Myopic Lasik

TRUE K : 38.31 / Corr. -5.28 D

ARTIS T PL E TORICA -aA Special TORICA -aA Special TORICA -aA Special

LF = 2.270 TK LF = 2.540	22.97	LF = 1.726 TK LF = 2.000	21.99	LF = 1.726 TK LF = 2.000	21.99	LF = 1.726 TK LF = 2.000	21.99
IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)	IOL(D)	REF(D)
22.00	-0.30	21.00	-0.27	21.00	-0.27	21.00	-0.27
22.50	-0.66	21.50	-0.64	22.00	-1.01	22.00	-1.01
23.00	-1.02	22.50	-1.39	22.50	-1.39	22.50	-1.39
23.50	-1.39	23.00	-1.77	23.00	-1.77	23.00	-1.77
24.00	-1.76						

Selected Lens lens constants / Ideal IOL Diopter

The closest IOL diopter of the input target

When selected IOL REF diopter, next button is available


Input target diopter after surgery

Input SIA (Surgically induced astigmatism)

Input IL (Incision on axis)

IOL : Comparison IOL Formula

IOL Calculation



* Wang, L., Shirayama, M., Ma, X.J., Kohnen, T., Koch, D.D. Optimizing intraocular lens power calculations in eyes with axial lengths above 25.0 mm. *J Cataract Refract Surg.* 2011;37:2018-2027.
 Abulafia, A., Barrett, G.D., Rotenberg, M., Kleinmann, G., Levy, A., Reitblat, O., Koch, D.D., Wang, L., Assia, E.I. Intraocular lens power calculation for eyes with an axial length greater than 26.0 mm: Comparison of formulas and methods. *J Cataract Refract Surg.* 2015;41:548-556.

* Scientifically the W-K adjustment is controversial since it depends on the IOL design whether the adjustment is required or not.

** Care has to be taken in eyes with very long AL. Special constants are needed for IOLs that change from symmetrical to asymmetrical geometry at certain power thresholds. This can lead to significant principal plane shifts and as a result to different ELP positions.

IOL : Comparison IOL Formula

IOL Calculation



Formula	ME±SD (D)	Mean AE	Median AE	Eyes within ±0.50 D prediction error (%)
Short axial length (n=89)				
Haigis	-0.080±0.428	0.344	0.280	75.3
Holladay 1	-0.097±0.421	0.334	0.295	77.5
SRK/T	-0.028±0.446	0.352	0.320	71.9
Barrett universal II	0.215±0.439	0.398	0.340	66.3
Medium axial length (n=432)				
Haigis	-0.126±0.439	0.360	0.285	73.6
Holladay 1	0.005±0.416	0.326	0.270	78.9
SRK/T	0.057±0.446	0.350	0.280	75.7
Barrett universal II	0.184±0.392	0.337	0.278	76.6
Long axial length (n=69)				
Haigis	-0.093±0.458	0.365	0.340	68.1
Holladay 1	0.306±0.521	0.500	0.410	56.5
SRK/T	0.064±0.464	0.386	0.350	69.6
Barrett universal II	0.079±0.406	0.327	0.300	78.3
Low power (n=84)				
Haigis	-0.020±0.394	0.318	0.288	77.4
Holladay 1	0.259±0.511	0.454	0.370	63.1
SRK/T	0.055±0.490	0.400	0.358	67.9
Barrett universal II	0.126±0.372	0.308	0.280	78.6
Medium power (n=436)				
Haigis	-0.133±0.478	0.362	0.283	73.6
Holladay 1	-0.013±0.412	0.323	0.270	78.9
SRK/T	0.025±0.437	0.342	0.280	76.6
Barrett universal II	0.164±0.391	0.332	0.278	77.3
High power (n=70)				
Haigis	-0.123±0.486	0.388	0.298	65.7
Holladay 1	-0.021±0.462	0.371	0.310	74.3
SRK/T	0.157±0.457	0.378	0.340	68.6
Barrett universal II	0.310±0.475	0.471	0.418	58.6

Holladay is suitable for Short AL

Barrett U II is suitable for Long AL

Barrett U II is suitable for Low IOL Power

Holladay is suitable for High IOL Power

ME, mean prediction error; SD, standard deviation of mean prediction error; AE, absolute prediction error; D, diopter

IOL : Comparison IOL Formula

IOL Calculation



60 Eyes:	True K TK	True K	Haigis L	Haigis TK	Shammas
Mean Error	-0.04	-0.17	-0.45	-0.12	-0.36
STDev	0.51	0.57	0.65	0.67	0.62
MAE	0.41	0.47	0.67	0.56	0.59
Med AE	0.34	0.37	0.61	0.50	0.57
</= % 0.5 D	70.0%	63.3%	31.7%	50.0%	46.7%
</= % 0.75 D	86.7%	85.0%	60.0%	71.7%	66.7%
</= % 1.00 D	96.7%	88.3%	83.3%	80.0%	86.7%

Table 5. Percentage of eyes within ± 0.50 D and ± 1.00 D from the target refraction in Groups A and B.

Formula/Method	Percentage			
	Group A (n = 58)		Group B (n = 30)	
	Within ± 0.50 D	Within ± 1.00 D	Within 0.50 D	Within ± 1.00 D
Adjusted Atlas	60.3	87.9	—	—
Masket	60.3	84.5	—	—
Modified Masket	53.4	86.2	—	—
Wang/Koch/Maloney	43.1	81.0	—	—
Shammas	55.2	82.8	50.0	80.0
Haigis-L	48.3	81.0	46.7	76.7
True-K	67.2	94.8	63.3	80.0
Average	58.6	91.4	46.7	83.3

Group A -
Pre/Post REF Data exist
Group B -
No history Pre/Post REF Data

Field Test : Israel



Field test

1) Measured Data

Name	Side	HBM-1						Lenstar-900					
		AL	CCT	ACD	LT	K1	K2	AL	CCT	ACD	LT	K1	K2
A**	R	22.36	563	2.729	4.08	46.07	46.83	22.35	550	3.22	4.04	45.81	46.57
	L	22.32	538	2.766	4.06	45.97	46.37	22.25	551	3.25	4.02	45.79	46.27
R**	R	23.36	518	3.12	4.26	45.15	46	23.35	527	3.55	4.23	44.93	45.75
	L	23.32	553	3.148	4.245	45.14	45.76	23.39	533	3.57	4.25	45.33	45.87
P**	R	22.37	557	3.5	3.63	45.86	46.9	22.36	548	3.96	3.63	45.82	47.16
	L	22.4	561	3.56	3.58	46.11	47.24	22.35	549	4	3.64	45.97	47.38
Y**	R	22.57	527	3.21	3.83	44.62	45.86	22.53	527	3.64	3.83	44.86	45.81
	L	22.81	523	3.27	3.83	44.82	45.75	22.71	526	3.68	3.82	44.84	45.67
S**	R	22.16	506	2.93	3.73	45.19	46.86	22.2	511	3.32	3.81	45.06	46.72
	L	22.09	526	2.97	3.82	44.62	47	21.96	511	3.36	3.82	45.06	47.69
G**	R	24.31	546	2.85	4.28	45.44	45.58	24.39	529	3.28	4.34	45.04	45.45
	L	24.08	545	4.2	0.89	44.52	46.44	24.16	535	4.65	0.83	45.14	45.74

2) Lens prescription

HBM			Holladay	Hoffer Q	Barrett	Lenstar-900		
22.0(-0.19)	22.0(-0.24)	22.0(-0.32)	0	0	0	22.0(-0.25)	22.0(-0.3)	22.0(-0.41)
21.5(-0.17)	21.5(-0.25)	21.5(-0.3)	-0.5	0	0	22.0(-0.37)	21.5(-0.1)	21.5(-0.18)
22.5(-0.08)	22.5(-0.08)	23.0(-0.16)	-0.5	-0.5	0	23.0(-0.39)	23.0(-0.39)	23.0(-0.16)
22.0(-0.36)	22.0(-0.4)	22.0(-0.09)	0	0	-0.5	22.0(-0.1)	22.0(-0.13)	22.5(-0.22)
23.5(-0.38)	23.5(-0.36)	23.0(-0.25)	0	0	0	23.5(-0.35)	23.5(-0.33)	23.0(-0.27)
24.0(-0.36)	24.0(-0.32)	23.5(-0.14)	0.5	0.5	0	23.5(-0.12)	23.5(-0.08)	23.5(-0.36)
16.5(-0.25)	16.0(-0.24)	16.5(-0.27)	0	0	0	16.5(-0.18)	16.0(-0.16)	16.5(-0.23)
17.0(-0.14)	17.0(-0.39)	17.0(-0.26)	0	0.5	---	17(-0.21)	16.5(-0.17)	??

Field Test : India



Field test

Comparison Of Biometry Measures Between HBM-1 And Lenstar-900

Parameter	HBM-1 Mean SD	Lenstar Mean SD	Diff	95% LoA
Axial Length	24.04±1.40	24.07±1.39	-0.03	`-0.08 to + 0.22
CCT	537.88±29.97	533.85±29.34	4.02	`-18.4 to +9.46
ACD	3.32±0.45	3.44±0.55	-0.12	`-0.38 to + 0.64
LT	4.32±0.45	4.13±0.47	0.18	`-0.83 to + 0.47
Flat Meridian(K1)	43.47±2.53	43.16±2.39	0.3	`-1.57 to + 0.95
Steep Meridian(K2)	44.28±2.59	44.21±2.65	0.06	`-0.95 to + 0.82
White to White	11.87±0.45	11.78±0.60	0.09	`-1.04 to + 0.87

Field Test : Romania

Field test

Comparison Of Biometry Measures Between HBM-1 And PentaCam-AXL				
Parameter	HBM-1 Mean SD	PentaCam Mean SD	Diff	95% LoA
Axial Length	23.56±0.85	23.52±0.88	0.04	`-0.04 to + 0.13
CCT	533±43.95	532±33.67	1.5	`-18.4 to +9.46
ACD	3.25±0.23	3.14±0.24	0.11	`-0.06 to + 0.15
Flat Meridian(K1)	43.93±1.27	43.58±1.21	0.35	`-0.12 to + 0.80
Steep Meridian(K2)	44.69±1.53	44.19±1.41	0.5	`-0.24 to + 1.23

Field Test : Romania

Field test

Comparison Of Biometry Measures Between HBM-1 And Optopol biometer				
Parameter	HBM-1 Mean SD	Optopol Mean SD	Diff	95% LoA
Axial Length	23.56±0.85	23.52±0.88	0.04	`-0.08 to + 0.16
CCT	533±43.95	548±35.67	14.6	`-42.7 to +14.3
ACD	3.25±0.23	3.16±0.24	0.09	`-0.02 to + 0.16
LT	4.56±0.34	4.62±0.36	-0.06	`-0.18 to + 0.07

Myopia management

Myopia



HBM-1

MYOPIA MANAGEMENT

Myopia Management 00001

OD OS

Atropine 0.01% Ortho-K 1.5Diopter

Atropine 0.02% Atropine 0.02% + Ortho-...

Atropine 0.02% + Otrho-...

AL Axial Length Graph

SPH SPH Graph

KER Keratometry Graph

ORTHO-K Ortho-K Comparison Topography

OD / OS Comparison OD / OS Information

DATA Edit / Add / Delete Data



AsianMale



Myopia management

Myopia



HBM-1

MYOPIA MANAGEMENT


Myopia Management 00001

- OD OS
- AL
- SPH
- KER
- ORTHO-K
- Axial Length(mm)
- DATA

Add Comment

- Atropine 0.01%
- Ortho-K 1.5Diopter
- Atropine 0.02%
- Atropine 0.02% + Ortho-...
- Atropine 0.02% + Otrho-...

Show Comment List



Years(Y)

AsianMale

Axial Length Trend by Race

Trend ON / OFF

Myopia management

Myopia



MYOPIA MANAGEMENT Myopia Management 00001

OD OS

AL

SPH

KER

ORTHO-K

OD / OS

DATA

Start. D 20100930 End. D 20120731

Start Date(YYYYMMDD) End Date(YYYYMMDD)

Comment Atropine 0.01%

Input Comment

Delete current comment

DELETE

SAVE

Save current comment

Age	Axial Length (mm)
4.0	21.00
5.0	21.20
6.0	21.70
7.0	22.20
8.0	22.70
9.0	22.50
10.0	22.80
11.0	23.00
12.0	23.00
13.0	23.30
14.0	24.20
16.0	27.30

AsianMale

Myopia management

Myopia



HBM-1

MYOPIA MANAGEMENT

Myopia Management 00001

Graph Scale

Start Year > 2.0 End Year

Start Year: 4 End Year: 16

Show Value: ON OFF

Value overlay ON/OFF

Cancel OK

The difference between the start year and the end year must set at least 2.0

Showing Myopia Trend Percentile

OD / O

DATA



AsianFemale



Myopia management

Myopia



HBM-1
MYOPIA MANAGEMENT
Myopia Management 90001

OD OS

AL

SPH

KER

ORTHO-K

OD / OS

DATA

3 / 5 / 7, 2 / 4 / 6 meridian changeable

2024-02-13 10:05

Before Topography wearing Ortho-k Treatment

3mm Zone

K-AVG 8.70mm
Cyl -0.26 @ 146

5mm Zone

K-AVG 8.46mm
Cyl -0.80 @ 178

7mm Zone

K-AVG 8.05mm
Cyl -1.18 @ 174

DIFF. MAP

Map Scale

ISO **ISO / American / Basic**

Interval **interval value on map**

0.1 **0.2**

Map Type

Axial Tangential

2024-02-01 20:47

After Topography wearing Ortho-k Treatment

3mm Zone

K-AVG 7.87mm
Cyl -0.51 @ 11

5mm Zone

K-AVG 7.90mm
Cyl -0.64 @ 9

7mm Zone

K-AVG 8.08mm
Cyl -0.68 @ 14

Myopia management




Myopia



- OD OS
- AL
- SPH
- KER
- ORTHO-K
- OD / OS
- DATA

2023-09-19

2023-08-08



Myopia management

Myopia



HBM-1
MYOPIA MANAGEMENT
Myopia Management S 00001

OD

OS

AL

SPH

KER

ORTHO-K

OD / OS

DATA

Raw Data		Add	Delete			
		Add Data	Delete Data			
	DATE	RX SE(D)	AL (mm)	ACD (mm)	LT (mm)	KER (D)
✓	19 - 09 - 2023	-5.00	25.90	3.47	4.00	45.00
✓	19 - 09 - 2023		26.45	3.74	3.82	42.80
✓	08 - 08 - 2023	-4.75	27.50	3.00	3.00	45.00
✓	08 - 08 - 2022	-4.75	27.30	3.00	3.00	44.00
✓	08 - 08 - 2021	-4.50				
✓	08 - 08 - 2020	-3.50	24.20	3.00	3.00	38.00
✓	08 - 08 - 2019	-1.75	23.30	3.00	3.00	38.00

Add

Delete

Settings

Print

←

1 / 3

→

Myopia management

Myopia

Edit Exam

Date

YY 2024 MM 02 DD 13

SPH

OD OS Load Setting

Biometry

AL 26.50 ACD LT

KER

Cancel OK

Must be Input

Load Sph
Data from
Huvitz RK

Port Setting
to connect
Huvitz RK

At least one
data must be
input

Myopia management

Myopia



Setup

REF

KER

AUTO START

COMMUNICATION

PRINT

DISPLAY

PATIENT NUMBER

ETC

I BPS(COM1) Setting BPS (9600 / 57600 / 115200)

9600	57600	115200
------	-------	--------

I RS232(COM1) Choose anything except 'OFF'

Off	V1	V2	Ext
-----	----	----	-----

I MODE(COM1)

Mate	Nidek	Topcon
------	-------	--------

I HLM PRINT

Off	On
-----	----

< 1 2 >

Set a Huvitz RK

HuvitzRK Connect Setting

Port

COM5

Bps


115200

Cancel OK


Set a HBM-1

Myopia management

Myopia



Press Load



HRK Press Transfer
And SPH Data Received

Myopia management

Myopia



MYOPIA Management 00001

Print AL / KER / REF Graph

Graph: AL KER REF

Graph Option: Start Year 2.0 ~ End Year 18.0 Myopia Trend: Asian Male

Editable Start year / end year

OrthoK: ON OFF

Select AL Trend (OFF / ON)

Before: 2024-02-13 10:05 After: 2024-02-01 20:47

Select Before Topo / After Topo

Compare: ON OFF

Before: 2023-09-19 After: 2023-08-08

Select OD / OS Data

Cancel OK

LT (mm)	KER (D)
4.00	45.00
3.82	42.80
3.00	45.00
3.00	44.00
3.00	38.00

Myopia management


Myopia




Huvitz Name: Patient_1_myopia ID: 00001 Gender: Male Exam Date: 08-08-2023
 Date of Birth: 2006-07-31 Physician: Exam Time: Operator:

MYOPIA MANAGEMENT REPORT (AXIAL LENGTH)

OD



OS



Data Graph + Trend

Myopia Treatment

- A Atropine 0.01%
- B Atropine 0.02%
- C Atropine 0.02% + Ortho-K 2.0D
- D Ortho-K 1.5Diopter
- E Atropine 0.02% + Ortho-K toric
- F

Myopia Treatment

- A Atropine 0.01%
- B Atropine 0.015%
- C Atropine 0.02% + Ortho-K 1D
- D Lasik Surgery
- E
- F

Comment

Date	AL(mm)	SE(D)	KER(D)
19-09-2023	25.90 (-0.50)	-5.00 (0.25)	45.00 (+4.25)
08-08-2023	26.40 (-0.90)	-4.75 (0.00)	40.75 (-3.25)
08-08-2022	27.30 (+3.10)	-4.75 (1.25)	44.00 (+6.00)
08-08-2020	24.20 (+0.90)	3.50 (-1.75)	38.00 (0.00)
08-08-2019	23.30 (+0.30)	-1.75 (0.25)	38.00 (+2.50)
08-08-2018	23.00 (0.00)	-1.50 (-0.25)	38.80 (-0.50)

Date	AL(mm)	SE(D)	KER(D)
08-08-2023	25.30 (-1.48)	-4.75 (2.00)	40.27 (2.87)
08-08-2022	26.78 (+0.82)	-2.75 (-1.08)	43.14 (+1.16)
08-08-2021	25.96 (-1.35)	-1.67 (+1.33)	41.98 (+0.53)
08-08-2020	27.31 (-0.04)	3.00 (-0.80)	41.45 (-1.28)
08-08-2019	27.35 (+0.35)	-2.20 (0.95)	42.73 (+3.93)
08-08-2018	27.00 (+0.30)	1.25 (+0.25)	38.84 (-3.84)

6 recent data

Comments | Signature | Device / SW info.

User Setting



User Setup

SETUP

System

Language English

Patient

Device Name Reset

Measure

Sleep Time Off 5 Min 10 Min 30 Min

Connectivity

Touch Keyboard On Off

Report

Login Page On Off

IOL

Save Mode Light Raw

Information

Contact Lens Import Export

Dicom

Light - Total 1 Measure -> 20MB
But, Not supported DCM
(But it can be done before saving)

Raw - Total 1 Measure -> 60MB

Default



1 / 1



Cancel

OK

User Setting (Import Contact Lens)

User Setup

HBM_ContactLensFitting_Database.csv - Excel

데이터가 손실될 수 있음 이 통합 문서를 심표로 구분된 형식(.csv)으로 저장하면 일부 기능이 손실될 수 있습니다. 기능을 유지하려면 Excel 파일 형식으로 저장하세요. 다시 표시 안 함 다른 이름으로 저장...

	A	B	C	D	E	F	G	H	I	J	K	L	M	N
1	Manufacturer	Model	Type	Base_r0	Base_Diameter	Base_Ecc	Periphery_r1	Periphery_b1	Periphery_r2	Periphery_b2	Periphery_r3	Periphery_b3	Periphery_r4	Periphery_b4
2	Appenz.	EX AS	Normal	8.45	9.8	0.6	11.95	0.4	0	0	0	0	0	0
3	Appenz.	EX MK	Normal	8.45	9.8	0.6	8.93	0.96	11.95	0.4	0	0	0	0
4	Falco	FAE	Normal	8.45	10	0.6	12.05	0.5	0	0	0	0	0	0
5	Falco	FAS	Normal	8.45	10	0.6	12.05	0.5	0	0	0	0	0	0
6	Galifa	ModulaA	Normal	8.45	9.8	0.6	11.25	0.4	10.7	0	0	0	0	0
7	Galifa	ModulaM	Normal	8.45	9.8	0.6	9.29	0.57	11.25	0.5	0	0	0	0
8	Hetych	Aktiv Comf.	Normal	8.45	10	0.6	9.05	0.65	12	0.2	0	0	0	0
9	Hetych	Aktiv Top	Normal	8.45	10	0.6	12	0.4	12	0	0	0	0	0
10	Hetych	Bifo Aktiv	Normal	8.45	10	0.6	9.05	0.6	12	0.2	0	0	0	0
11	JenaLens	ABOGDT	Normal	8.5	9.8	0.45	12	0.4	12	0	0	0	0	0
12	JenaLens	ABOGVTP	Normal	8.5	9.8	0.45	12	0.4	12	0	0	0	0	0
13	JenaLens	ABOGTpri	Normal	8.5	9.8	0.45	12	0.4	12	0	0	0	0	0
14	JenaLens	Asph	Normal	8.45	9.8	0.6	12	0.4	12	0	0	0	0	0
15	JenaLens	Asph-Kera	Normal	8.25	9.8	0.9	12	0.4	12	0	0	0	0	0
16	JenaLens	AsphPri	Normal	8.45	9.8	0.6	12	0.4	12	0	0	0	0	0
17	JenaLens	Asph VTP	Normal	8.45	9.8	0.6	12	0.4	12	0	0	0	0	0
18	JenaLens	Jeclips	Normal	8.3	9.8	0.8	0	0	0	0	0	0	0	0
19	JenaLens	Jecl. RT/BT	Normal	8.3	9.8	0.8	0	0	0	0	0	0	0	0
20	JenaLens	Jecl. VT	Normal	8.3	9.8	0.8	0	0	0	0	0	0	0	0
21	MPG&E	compact AS	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
22	MPG&E	compact ASAB	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
23	MPG&E	compact ASVPT	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
24	MPG&E	compact CX	Normal	8.5	9.8	0.45	0	0	0	0	0	0	0	0
25	MPG&E	compact CXVP	Normal	8.5	9.8	0.45	0	0	0	0	0	0	0	0
26	MPG&E	compact CXV	Normal	8.5	10	0.45	0	0	0	0	0	0	0	0
27	MPG&E	compact EZ	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
28	MPG&E	compact EZV	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
29	MPG&E	compact EZM	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
30	MPG&E	compact EZMVT	Normal	8.45	10	0.6	0	0	0	0	0	0	0	0
31	PRECI	AZ	Normal	8.6	9.8	0.3	10	0.2	10	0.2	0	0	0	0
32	PRECI	AZTE	Normal	8.6	9.8	0.3	10	0.2	10	0.2	0	0	0	0
33	PRECI	BA7	Normal	8.6	9.8	0.3	10	0.2	10	0.2	0	0	0	0

User Setting



User Setup

SETUP

System

Patient List Size

50

100

150

200

500

Patient

Today List

On

Off

Measure

PID Prefix

Reset

Connectivity

PID Postfix

Reset

Report

PID Number Length

5

6

7

8

IOL

Data Format

YMD

MDY

DMY

Information

Staff Management

Physicians

Set the name of a doctor which is used to define IOL setting



1 / 1



Cancel

OK

User Setting



User Setup

SETUP

System

Auto Tracking

On

Off

Auto Retry

On

Off

Patient

Auto Shoot

On

Off

Measure automatically until 3 times
If data is not successful

Measure

Auto Shoot Detail

KER

AL

ANT

Connectivity

Live Signal

On

Off

Map Type

REF Power

Elevation

Report

Measurement Scale

Diopter

Millimeter

Axial

Tangential

IOL

Keratometry
Display Area

2-4-6mm

3-5-7mm

Meridian Unit

Information

Cylinder Notation

Negative

Positive

Dicom

Default



1 / 3



Cancel

OK

User Setting



User Setup

SETUP

System

Asphericity

e

SF

p

Q

Patient

Measurement Mode

KER

AL

ANT

Types of Asphericity

Measure

Axial Length Offset

0

Axial Length Std

0.045

Connectivity

K1 Offset

0

Evaluation Red

2

AL settings

Report

K2 Offset

0

Evaluation Yellow

3

IOL

Evaluation Green

4

Information

Dicom

Default



2 / 3



Cancel

OK

User Setting



User Setup

SETUP

System

CCT Offset

0

CCT Std

0.015

Patient

ACD Offset

0

ACD Std

0.05

Measure

LT Offset

0

LT Std

0.06

Connectivity

Evaluation Red

3

ANT settings

Report

Evaluation Yellow

4

IOL

Evaluation Green

5

Information

Dicom

Default



3 / 3



Cancel

OK

User Setting

User Setup



SETUP

System

Auto Data Transfer

On

Off

Transfer data when saving

Patient

HIIS-1 Server IP

hocr.huvitz.com

Reset

Measure

HIIS-1 Server Port

8080

Reset

Connectivity

Report

IOL

Information



1 / 1



Cancel

OK

User Setting



User Setup

SETUP

System

Report Logo

C:\Users\Huvitz\Desktop\ta\pupil2.png

...

Reset

Patient

Organization

Test

Reset

Measure

Auto Export

On

Off

Connectivity

Auto Export Folder

D:\HBM_Export_Report

...

Reset

Report

Export List

Biometry

Topography

IOL

IOL

Export TopoMap

Axial

Tangential

Elevation

Ref Power

Information



1 / 1



Cancel

OK

User Setting



User Setup

SETUP

System

Physician 1

IOL Lens Editor Open

IOL Default

Patient

Choose a physician from 1 .. 6

Open IOL Editor

Factory initialization

IOL Calculation

Measure

Target Refraction 0.00

Connectivity

Manufacturer AMO AMO AMO AMO

Report

Model Tecnis 1 ZCB... Tecnis 1 ZCB... Tecnis 1 ZCB... Tecnis 1 ZCB...

IOL

Formula SRK2 SRK2 SRK2 SRK2

Information

Dicom

Default



1 / 3



Cancel

OK

User Setting

User Setup



SETUP

System

Physician

1

IOL Lens Editor

Open

Patient

Post Refractive IOL

Choose a physician
from 1 .. 6

Open IOL Editor

Measure

Target Refraction

0.00

Connectivity

Manufacturer

AMO

AMO

AMO

AMO

Report

Model

Tecnis 1 ZCB...

Tecnis 1 ZCB...

Tecnis 1 ZCB...

Tecnis 1 ZCB...

IOL

Formula

Shammas N...

Shammas N...

Camellin-Cal...

Camellin-Cal...

Information



2 / 3



Cancel

OK

User Setting

User Setup

SETUP

System

Physician

1

IOL Lens Editor

Open

Patient

Toric

Choose a physician
from 1 .. 6

Open IOL Editor

Measure

Target Refraction

0.00

Connectivity

Manufacturer

HOYA

HOYA

HOYA

HOYA

Report

Model

iSert Toric 351

iSert Toric 351

iSert Toric 351

iSert Toric 351

IOL

Formula

Haigis

Barrett Univ...

Holladay

SRK/T

Information



3 / 3



Cancel

OK

User Setting



User Setup

SETUP

System

S/W Version v1.1.0-Pre(2023_02_13_1000)

Patient

Organization Select NewUser Password Delete

Measure

~01-01-0001 Extend NewAdmin

Connectivity

License Barrett

Report

Serial Number 1BM00023E0006

IOL

Change Directory Backup Data

Information

Storage D:\Measurement Data\ ... Backup



Dicom

Default



1 / 1



Cancel

OK

User Setting (Change Directory)

User Setup



Setting Directory

D:\Measurement Data\ Factory initialization

Patient Data	OK	Init
IOL Lens Data	OK	Init
Contact Lens Data	OK	Init

Dir Close


Change Dir

이름	수정한 날짜	유형	크기
174	2024-02-19 오후 5:19	파일 폴더	
175	2024-01-15 오전 11:48	파일 폴더	
177	2024-01-31 오후 3:23	파일 폴더	
179	2024-01-31 오후 3:23	파일 폴더	
ContactLens.db	2024-02-16 오후 3:14	Data Base File	32KB
HBM_DB2.db	2024-02-19 오후 1:16	Data Base File	134KB
iolc.db	2024-01-18 오후 4:30	Data Base File	71KB

Caution !
When you change the path,
database files must have in
that path

User Setting (Change Directory)

User Setup



Setting Directory

D:\Measurement Data_\

Patient Data	OK	Init
IOL Lens Data	Error	Init
Contact Lens Data	Error	Init

Dir Close

If database file is damaged, click 'Init' button to initialize it. But this data will be in a factory initialization state

Setting Directory

D:\Measurement Data_\

Patient Data	OK	Init
IOL Lens Data	OK	Init
Contact Lens Data	OK	Init

Dir Close

User Setting (Change Directory)

User Setup



Backup_2024_01_10_07_03_11_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_03_16_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_12_46_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_13_00_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_13_33_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_13_44_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_07_17_22_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_11_05_26_05_HBM_DB2.db	2024-01-10 오후 2:42	Data Base File	134KB
Backup_2024_01_10_15_18_ContactLens.db	2024-01-10 오후 3:18	Data Base File	0KB
Backup_2024_01_10_15_18_iolc.db	2024-01-10 오후 3:18	Data Base File	0KB
Backup_2024_01_11_02_54_41_HBM_DB2.db	2024-01-11 오전 11:36	Data Base File	129KB
Backup_2024_01_11_03_00_01_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_04_35_05_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_04_59_01_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_05_01_10_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_05_23_13_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_05_28_17_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_05_28_48_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_05_29_41_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_06_06_22_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_06_06_53_HBM_DB2.db	2024-01-11 오전 11:48	Data Base File	134KB
Backup_2024_01_11_06_20_50_HBM_DB2.db	2024-01-11 오후 3:19	Data Base File	0KB
Backup_2024_01_11_06_20_51_iolc.db	2024-01-11 오후 3:19	Data Base File	0KB
Backup_2024_01_11_06_20_52_ContactLens.db	2024-01-11 오후 3:19	Data Base File	0KB
Backup_2024_01_11_06_25_45_iolc.db	2024-01-11 오후 3:25	Data Base File	0KB
Backup_2024_02_19_08_37_04_iolc.db	2024-02-19 오후 5:37	Data Base File	0KB
Backup_2024_02_19_08_37_05_ContactLens.db	2024-02-19 오후 5:37	Data Base File	0KB

When user Pressed 'Init',
database will be backup
(C:\HBM-1\DatabaseBackup)

User Setting (Data Backup)

User Setup

Backup Database

Source Directory
D:¥Measurement Data¥
Space Used : 1.75GB

Destination Directory
C:¥HBMBACKUPTEST¥
Total Size : 459.42GB Free Size : 215.78GB

0.0%

Start
Cancel
Close

Amount of data to be backup

Select Backup Directory

Selected Backup Directory

Selected Backup Directory Information

If Free Size smaller than Data size, Can't Backup

During the initialize patient database, you can cancel initialization

User Setting (Data Backup)

User Setup

Backup Database

Source Directory
D:¥Measurement Data¥
Space Used : 1.85GB

Destination Directory
C:¥HBMBACKUPTEST¥
Total Size : 459.42GB Free Size : 215.70GB

Progressing
48.74%

IOL Editor



User Setup

iolcon.org IOL Con - Intraocular Lenses

Steinbeis Vision Research **IOL Con**

Show All Lenses Search For Lenses Add Clinical Result Add Lens Log In

Download Print Subscribe 652 selected lenses

Constants to download:

- Biometer: all
- Ethnicity: all
- Institution: all
- Surgeon: all
- File Version: 1.7
- Manufacturer / ULIB
- IOL Con

Our Optimized Constants³

Haigis (a ₀ / a ₁ / a ₂)	Hoffer Q (pACD)	Holladay 1 (SF)	SRK/T (A)	Castrop (C / H / R)
0.8751 0.2659 0.1416	5.6	1.809	119.033	
0.8751 0.2659 0.1416	5.6	1.809	119.033	
0.8751 0.2659 0.1416	5.6	1.809	119.033	

Manufacturer / ULIB - Provide Constants Manufacturer IOL Con - Provide iolcon.org optimized constants

LPCM Help Contact Imprint Terms & Conditions Data Privacy Statement

IOL Editor

User Setup

IOL Lens Edit

Keyboard

Select DB :

Spherical

History Log

Manufacture	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (pA)	Holladay (SF)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin Cal	Shammas (A)
1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Q	118	118.3	118.1	5.01	1.25	0.274	0.458	0.116	1.412	118	118
1stQ	Basis Z	118	118.3	118.1	5.01	1.25	-0.07	0.21	0.163	1.412	118	118
1stQ	Basis Z h	118	119.2	118.9	5.46	1.7	1.32	0.4	0.1	1.831	118	118
AMO	808C	118	119.6	119.1	5.65	1.89	1.5	0.4	0.1	1.936	118	118
AMO	CeeOn 9	118.3	119.1	118.7	5.42	1.65	0.568	0.224	0.152	1.726	118.3	118.3
AMO	ClariFlex	118	118.6	118.3	5.14	1.37	0.92	0.4	0.1	1.517	118	118
AMO	PS60 AN	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7
AMO	ReZoom	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4
AMO	SA40 Av	118	118.1	117.9	4.9	1.14	0.65	0.4	0.1	1.207	118	118

Add New Lens

Delete Lens

Import
(XML)

Import
Sph. DB
(CSV)

Export
Sph. DB
(CSV)

Save

Close

IOL Editor



User Setup

Import

Reference Type :

IOL Con

Constraint Type :

Norminal/ULIB

IOL Con Optimized

Load Data

Load USB

<input type="checkbox"/>	Manufactu	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (p)	Holladay (s)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin C	Shammas (
<input type="checkbox"/>	1stQ	611HPS	0	0	118.3	5.15	1.39	0.95	0.4	0.1	1.52	0	0
<input type="checkbox"/>	1stQ	B1AB00	0	0	118.9	5.46	1.7	1.32	0.4	0.1	1.9	0	0
<input type="checkbox"/>	1stQ	B1ABY0	0	0	118.9	5.46	1.7	1.32	0.4	0.1	1.9	0	0
<input type="checkbox"/>	1stQ	B1ADYC	0	0	118.9	5.46	1.7	1.32	0.4	0.1	1.9	0	0
<input type="checkbox"/>	1stQ	B1AP00	0	0	118.9	5.46	1.67	1.243	0.4	0.1	0	0	0
<input type="checkbox"/>	1stQ	B1ADY0	0	0	118.9	5.46	1.67	1.243	0.4	0.1	0	0	0

Import

Cancel

IOL Editor

IOL Lens Edit

Keyboard

Select DB :

Spherical

History Log

Manufacture	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (pA)	Holladay (SF)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin Cal	Shammas (A)
1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Q	118	118.3	118.1	5.01	1.25	0.274	0.458	0.116	1.412	118	118
1stQ	Basis Z	118	118.3	118.1	5.01	1.25	-0.07	0.21	0.163	1.412	118	118
1stQ	Basis Z h	118	119.2	118.9	5.46	1.7	1.32	0.4	0.1	1.831	118	118
AMO	808C	118	119.6	119.1	5.65	1.89	1.5	0.4	0.1	1.936	118	118
AMO	CeeOn 9	118.3	119.1	118.7	5.42	1.65	0.568	0.224	0.152	1.726	118.3	118.3
AMO	ClariFlex	118	118.6	118.3	5.14	1.37	0.92	0.4	0.1	1.517	118	118
AMO	PS60 AN	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7
AMO	ReZoom	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4
AMO	SA40 Av	118	118.1	117.9	4.9	1.14	0.65	0.4	0.1	1.207	118	118

Add New Lens

Delete Lens

Import
(XML)

Import
Sph. DB
(CSV)

Export
Sph. DB
(CSV)

Save

Close

IOL Editor

IOL Lens Edit

Keyboard

Select DB :

Spherical

History Log

Manufacture	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (pA)	Holladay (SF)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin Cal	Shammas (A)
1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Q	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Z	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Z h	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
AMO	808C	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
AMO	CeeOn 9	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
AMO	ClariFlex	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
AMO	PS60 AN	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7
AMO	ReZoom	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4
AMO	SA40 Av	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9

Spherical DB Import (CSV)

Path : C:\Users\Huvitz\Desktop\IOL_Sph_20240226_142132.csv

Open File

Data read successful.

Import

Apply

Cancel

Add New Lens

Delete Lens

Import
(XML)

Import
Sph. DB
(CSV)

Export
Sph. DB
(CSV)

Save

Close

IOL Editor

IOL Lens Edit

Keyboard

Select DB :

Spherical

History Log

Manufacture	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (pA)	Holladay (SF)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin Cal	Shammas (A)
1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Q	118	118.3	118.1	5.01	1.25	0.274	0.458	0.116	1.412	118	118
1stQ	Basis Z	118	118.3	118.1	5.01	1.25	-0.07	0.21	0.163	1.412	118	118
1stQ	Basis Z h	118	119.2	118.9	5.46	1.7	1.32	0.4	0.1	1.831	118	118
AMO	808C	118	119.6	119.1	5.65	1.89	1.5	0.4	0.1	1.936	118	118
AMO	CeeOn 9	118.3	119.1	118.7	5.42	1.65	0.568	0.224	0.152	1.726	118.3	118.3
AMO	ClariFlex	118	118.6	118.3	5.14	1.37	0.92	0.4	0.1	1.517	118	118
AMO	PS60 AN	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7
AMO	ReZoom	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4
AMO	SA40 Av	118	118.1	117.9	4.9	1.14	0.65	0.4	0.1	1.207	118	118

Add New Lens

Delete Lens

Import
(XML)

Import
Sph. DB
(CSV)

Export
Sph. DB
(CSV)

Save

Close

IOL Editor

IOL Lens Edit

Keyboard

Select DB :

Spherical

History Log

Manufacture	Model	Manu (A)	SRK2 (A)	SRKT (A)	HofferQ (pA)	Holladay (SF)	Haigis (a0)	Haigis (a1)	Haigis (a2)	Barrett (LF)	Camellin Cal	Shammas (A)
1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9
1stQ	Basis Q	118	118.2	118.1	5.01	1.25	0.974	0.459	0.116	1.412	118	118
1stQ	Basis Z	118									118	118
1stQ	Basis Z h	118									118	118
AMO	808C	118									118	118
AMO	CeeOn 9	118									118.3	118.3
AMO	ClariFlex	118	118.6	118.3	5.14	1.37	0.92	0.4	0.1	1.517	118	118
AMO	PS60 AN	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7
AMO	ReZoom	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4
AMO	SA40 Av	118	118.1	117.9	4.9	1.14	0.65	0.4	0.1	1.207	118	118

Spherical DB Export (CSV)

Path : ./IOL_Sph_20240226_142756.csv

Data export successful.

Open Folder

Export

Close

Add New Lens

Delete Lens

Import (XML)

Import Sph. DB (CSV)

Export Sph. DB (CSV)

Save

Close


Exported CSV file

User Setup

IOL_Sph_20240405_164802.csv - Microsoft Excel

A1	Manufacturer																										
	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z	
1	Manufact	Model	Manu_A	SRKII_A	SRK_T_A	HofferQ_p	Holladay_	Haigis_a0	Haigis_a1	Haigis_a2	Barrett_LF	Camellin_	Shammas_A														
2	1stQ	Basis K	118.9	118.6	118.3	5.15	1.37	0.95	0.4	0.1	1.517	118.9	118.9														
3	1stQ	Basis Q	118	118.3	118.1	5.01	1.25	0.274	0.458	0.116	1.412	118	118														
4	1stQ	Basis Z	118	118.3	118.1	5.01	1.25	-0.07	0.21	0.163	1.412	118	118														
5	1stQ	Basis Z hy	118	119.2	118.9	5.46	1.7	1.32	0.4	0.1	1.831	118	118														
6	AMO	808C	118	119.6	119.1	5.65	1.89	1.5	0.4	0.1	1.936	118	118														
7	AMO	CeeOn 91	118.3	119.1	118.7	5.42	1.65	0.568	0.224	0.152	1.726	118.3	118.3														
8	AMO	ClariFlex	118	118.6	118.3	5.14	1.37	0.92	0.4	0.1	1.517	118	118														
9	AMO	PS60 ANB	116.7	118.9	118.7	5.46	1.65	1.15	0.4	0.1	1.726	116.7	116.7														
10	AMO	ReZoom N	118.4	118.5	118.3	5.2	1.4	0.92	0.4	0.1	1.517	118.4	118.4														
11	AMO	SA40 Arra	118	118.1	117.9	4.9	1.14	0.65	0.4	0.1	1.307	118	118														
12	AMO	Sensar 1-p	118.4	119.3	119	5.57	1.79	-1.004	0.182	0.232	1.87	118.4	118.4														
13	AMO	Sensar AR	118.4	118.9	118.7	5.39	1.62	0.472	0.077	0.174	1.726	118.4	118.4														
14	AMO	Sensar AR	118.4	118.8	118.7	5.41	1.63	-2.42	0.157	0.288	1.726	118.4	118.4														
15	AMO	SI30 NB	117.4	118.6	118.5	5.34	1.58	1.23	0.4	0.1	1.622	117.4	117.4														
16	AMO	SI40 NB	118	118.6	118.4	5.19	1.42	-0.199	0.276	0.169	1.569	118	118														
17	AMO	SI55	118	119.1	118.7	5.28	1.56	1.1	0.4	0.1	1.726	118	118														
18	AMO	Tecnis 1 Z	118.8	119.6	119.3	5.8	2.02	-1.302	0.21	0.251	2.041	118.8	118.8														
19	AMO	Tecnis Z9C	119	119.5	119.2	5.71	1.93	-0.663	0.117	0.232	1.988	119	119														
20	AMO	Tecnis Z9C	119	119.4	119.2	5.71	1.94	1.5	0.4	0.1	1.988	119	119														
21	AMO	Tecnis Z9C	118.7	119	118.8	5.46	1.67	-1.201	0.102	0.246	1.779	118.7	118.7														
22	AMO	Tecnis ZA	119.1	119.4	119.1	5.61	1.84	-1.298	0.233	0.24	1.936	119.1	119.1														
23	AMO	Tecnis ZM	119	120.5	119.8	5.89	2.21	1.6	0.4	0.1	2.303	119	119														
24	AMO	Tecnis ZM	119.1	120	119.5	5.8	2.06	-1.75	0.242	0.266	2.146	119.1	119.1														
25	AMO	Tecnis ZM	118.8	119.7	119.5	5.88	2.1	1.68	0.4	0.1	2.146	118.8	118.8														
26	AMO	Tecnis ZM	118.8	119.7	119.7	6.14	2.35	1.93	0.4	0.1	2.25	118.8	118.8														
27	AMO	Verisyse 5i	116.8	116.8	116.9	4.34	0.54	-0.25	0.4	0.1	0.784	116.8	116.8														
28	AMO	ZFR00V Sy	119.3	0	119.211	5.629	1.882	1.456	0.4	0.1	0	0	0														
29	ARGONOF	EUROMA	118	118.5	118.2	5.09	1.33	0.87	0.4	0.1	1.465	118	118														
30	ARGONOF	TRICE	118	118.7	118.5	5.28	1.49	1.13	0.4	0.1	1.622	118	118														
31	AUROLAB	FH5600AS	118	117.9	117.8	4.92	1.12	0.68	0.4	0.1	1.255	118	118														
32	Aaren	Scientific	118.5	119.2	118.8	5.54	1.74	1.36	0.4	0.1	1.779	118.5	118.5														
33	Aaren	Scientific	118.3	119.1	118.7	5.37	1.63	1.22	0.4	0.1	1.726	118.3	118.3														
34	Aaren	Scientific	117.8	118.8	118.7	5.4	1.61	1.2	0.4	0.1	1.726	117.8	117.8														
35	Aaren	Scientific	118.2	119.4	119.1	5.69	1.89	1.45	0.4	0.1	1.936	118.2	118.2														
36	Aaren	Scientific	118.1	118.2	118	4.97	1.2	0.75	0.4	0.1	1.36	118.1	118.1														
37	Alcon	AcrySof M	118.4	118.8	118.7	5.46	1.64	1.34	0.4	0.1	1.726	118.4	118.4														

IOL Editor



IOL Lens Edit

Keyboard

Select DB :

Toric

History Log

Manufacturer

- Alcon
- AMO**
- HOYA
- Oculentis
- HumanOptics

Model

Tecnis ZCTx

Lens Factor

- Manu A 118.800
- SRK/T A 119.300
- HofferQ pACD 5.800
- Holladay SF 2.020
- Haigis A0 -1.302
- Haigis A1 0.210
- Haigis A2 0.251
- Barrett LF 2.041

Sphere Power Range

- Min 5
- Max 34
- Step 0.5

Cylinder Power Range

Lens	Toricity
Tecnis ZCT100	1
Tecnis ZCT150	1.5
Tecnis ZCT225	2.25
Tecnis ZCT300	3
Tecnis ZCT375	3.75

Add New Toric Lens

Edit Toric Lens

Delete Toric Lens

Close

IOL Editor

User Setup

Edit Toric Lens

Keyboard

Manufacturer and Model

Manufacturer

Model/Series

Cylinder Range

Sphere Power Range

Min Max Step

SubModel List

Lens	Toricity

Add

Delete

Lens Factor

Manu A

HofferQ PACD

Holladay SF

Haigis A0

Haigis A1

Haigis A2

Barrett LF

Save

Cancel

Exist Submodel & UnConstant Toricity Step

Not exist Submodel & Constant Toricity step

IOL Editor



User Setup

Edit Toric Lens

Keyboard

Manufacturer and Model

Manufacturer

Model/Series

Cylinder Range

Sphere Power Range

Min Max Step

SubModel List

Lens	Toricity
AlConSubmodelTest1	0.5
AlConSubmodelTest2	1.5
AlConSubmodelTest3	2.25
AlConSubmodelTest4	3
AlConSubmodelTest5	3.5

Lens Factor

Manu A

SRK/T A

HofferQ pACD

Holladay SF

Haigis A0

Haigis A1

Haigis A2

Barrett LF

IOL Editor



User Setup

Edit Toric Lens Keyboard

Manufacturer and Model

Manufacturer:

Model/Series:

Cylinder Range:

Sphere Power Range

Min: Max: Step:


Cylinder Power Range

Min: Max: Step:

Lens Factor

Manu A	<input type="text" value="0.0"/>
SRK/T A	<input type="text" value="0.0"/>
HofferQ pACD	<input type="text" value="0.0"/>
Holladay SF	<input type="text" value="0.0"/>
Haigis A0	<input type="text" value="0.0"/>
Haigis A1	<input type="text" value="0.0"/>
Haigis A2	<input type="text" value="0.0"/>
Barrett LF	<input type="text" value="0.0"/>

Q&A



Q & A

